



Electronic Temperature Instruments Ltd
Easting Close
Worthing
West Sussex
BN14 8HQ
Tel +44 01903 202151
Fax+44 01903 202445

MANUFACTURER'S CERTIFICATE OF CALIBRATION

Date Of Issue: Tuesday, 01 February 2011

Certificate Issued By: Debbie Board

Instrument Description: THERMA K THERMOMETER

Instrument Serial Number: D11040360

Ambient Temperature: 20°C±2°C Ambient Humidity: 40% ±10%

Reference Instrument UKAS Certificate Number: CN102783 & N1187858T

Results:	Test Temperature °F	Instrument Reading °F
	32	32.0
	212	212.0

Uncertainty of Calibration ± 0.4°C

This is to certify that the above Instrument has been calibrated against laboratory standards which are traceable via International Agreement, to all major National Standards, including the NPL and NIST.

Authorised by: John Carswell

Electronic Temperature Instruments Ltd


Manufacturer's Statement of Conformity

Date as overleaf

This Statement of Conformity confirms that the instrument referred to overleaf has been designed, manufactured, tested, inspected and stored in conformity with ETI's relevant procedures and specifications.

The above procedures have been undertaken in accordance with the Company's BS EN ISO 9001:2000 Quality System, as registered and approved by BSI Quality Assurance. Registered Certificate No. FM38039.

The relevant specifications will include all required standards in respect of EMC Directives, Health & Safety legislation, and also HACCP recommendations.



David Carter
Quality and Systems Manager

CONTINUED MEASUREMENT CONFIDENCE

The Manufacturer's Certificate of Calibration and the Statement of Conformity issued with your instrument are your guarantee that you have purchased a quality product that will provide you with years of trouble-free service. In the unlikely event of a problem, ETI's manufactured instruments carry a 24-month warranty against component failure or manufacturing defect. Other instruments have a 12-month warranty, and probes have a six-month warranty period.

The Manufacturer's Certificate of Calibration overleaf is not issued under ETI's accreditation to UKAS. However the Certificate is traceable to National Standards.

For continued confidence in the accuracy of measurement, and to assist in compliance with legislation, it is recommended that you obtain a Certificate of Calibration at 12-month intervals. ETI is able to provide UKAS Certificates from its Worthing based UKAS Laboratories.

The United Kingdom Accreditation Service (UKAS) is the UK national accreditation body responsible for assessing and accrediting the competence of organisations in the fields of measurement, testing, inspection and certification of systems, products and personnel. It operates under a Memorandum of Understanding with the Department of Trade and Industry, which has licensed UKAS to use accreditation marks that feature the Royal Crown, and to sub-license the use of these marks to UKAS accredited organisations.

Having achieved accreditation to UKAS, ETI's laboratories are inspected on a regular basis to ensure that our calibration methods, equipment and personnel are maintained to the highest standards.

**Full details of ETI's calibration service may be obtained from the sales office,
Tel: 01903 202151 · Fax: 01903 202445 · Website: www.etiltd.com**