

GUARANTEE

The thermometer is guaranteed for a period of one year from the date of purchase against mechanical and electrical manufacturing defects. There are no user serviceable parts inside the instrument. Any attempted repair by unauthorized persons voids the warranty.

Avoid keeping the thermometer too close to objects that continuously generate high heat for long periods (like a hot plate), which can cause overheating of the lens.

SPECIFICATIONS

Thermometer Measurement Range: -67 to +428°F / -55 to +220°C
 Ambient Operating Range: 14 to +122°F / -10 to +50°C
 Display Resolution: 0.1° between -19.9 to 199.9°, otherwise 1°
 Temperature Scale: °F / °C switchable
 Response Time: 1 second
 Accuracy (Tobj=15 to 35°C, Tamb=25°C)*: 1.1°F / ±0.6°C
 Accuracy (Tobj=-33 to 220°C, Tamb=23 ±3°C)*: ±2% of reading or 4°F / 2°C whichever is greater
 Distance/Target Ratio: 1:1 optics ratio
 Emissivity: Adjustable from 0.05 to 1.00 in .01 steps
 Battery Type: (1) CR2032 Lithium Cell, included
 Battery Life: Typically 40hr., minimum of 30hr
 Auto Shutoff: After 15 seconds
 Dimensions: 1.9 (W) x 2.9 (H) x 0.75 (D) inches
 48.5 (W) x 75.4 (H) x 19 (D) mm
 Weight: 1.5 oz / 42.3 g including battery

*IR accuracy can be affected by the emissivity setting, target spot size, rapid changes in ambient temperature.

07-12-13 IR-Mini © ThermoWorks, Inc. All rights reserved.

5



ThermoWorks

CERTIFICATE OF CONFORMANCE

The manufacturer of this instrument has implemented a quality assurance system under ISO 9001:2000 certified quality system and fully follow ISO GUM (Guide to the Expression of Uncertainty in Measurement) to evaluate the uncertainty of temperature and resistance standards, guarantee performance as below:

Calibration Temperature	Max Error	Target Accuracy	Target Stability
-30°C	±2.0°C	±0.3°C	0.3°C
0°C	±2.0°C	±0.4°C	0.1°C
100°C	±2.0°C	±0.4°C	0.1°C
200°C	±4.0°C	±0.5°C	0.1°C

Note: Assume the operation ambient temperature under 23 ±3°C

Furthermore, we certify that this infrared thermometer has been inspected and found to comply with published specifications. This device has been calibrated by temperature standards that are traceable to NIST (U.S. National Institute of Standards and Technology) or NML (National Measurement Laboratory) of Taiwan, and the calibration procedure corresponds with generally accepted regulations and standards.

ThermoWorks, Inc.
 1762 W. 20 S. #100
 Lindon, UT 84042
 www.thermoworks.com


6

Operating Instructions

The IR-Mini is a non-contact infrared thermometer designed to measure the surface temperature of various objects. Simply aim the thermometer at the target and press the SCAN button to display the surface temperature. The distance to target ratio is 1:1, therefore the thermometer should be positioned as close to the target as possible for best results.

Information is displayed on the LCD through these symbols:

- A. Temperature reading
- B. Minus sign (negative temperature)
- C. Battery status
- D. MINIMUM mode
- E. MAXIMUM mode
- F. LOCK mode
- G. Selected temperature scale (°C or °F)




°C OR °F MODE

To change the thermometer from °C to °F or from °F to °C, first turn the instrument on by pressing the SCAN button, then press the MODE button four times. The °C or °F symbol will flash. Press the SCAN button to change and confirm the new units.

MINIMUM OR MAXIMUM MODE

To utilize the thermometer's minimum or maximum mode, first turn the instrument on by pressing the SCAN button, then press the MODE button once for minimum or twice for maximum function. The MIN or MAX icon will flash, then press the SCAN button



7

to confirm the minimum or maximum mode. Keep pressing the SCAN button and the thermometer will display the minimum or maximum reading only.

LOCK MODE

The LOCK mode locks the thermometer in SCAN mode, and is particularly useful for continuous monitoring of temperatures. To utilize the thermometer's LOCK mode, first turn the instrument on by pressing the SCAN button, then press the MODE button three times for the LOCK mode function. The LOCK icon will flash. Then press the SCAN button to confirm the LOCK measurement mode. The thermometer will continuously display the temperature for up to 60 minutes or until the SCAN button is pressed.

EMISSION

The infrared thermometer is supplied with a default emissivity of 0.95. The emissivity of the thermometer can be changed from 0.05 (5E) to 1.00 (100E). Changes should only be carried out by experienced personnel. To change the emissivity, first turn the instrument on by pressing the SCAN button, then press the MODE button five times for emissivity function. The LCD display will flash 95E. Then press the SCAN button to adjust the emissivity value to the desired setting. Press the MODE button again to exit the set up screen. For information regarding emissivity and the emissivity of specific materials, please visit our website at www.thermoworks.com.

Note: non-contact infrared thermometers are not recommended for measuring the temperature of shiny or polished metals.

LCD ERROR MESSAGES

The thermometer incorporates visual diagnostic messages as follows:

Er2 Er2 is displayed when the thermometer is exposed to rapid changes in the ambient temperature.

2

Er3 Er3 is displayed when the ambient temperature exceeds 14°F / -10°C or 122°F / 50°C. The thermometer should be allowed plenty of time (minimum 30 minutes) to stabilize to the ambient room temperature.

Er For all other error messages it is necessary to reset the thermometer. To reset, turn the thermometer off, remove the battery and wait for a minimum of one minute. Reinsert the battery and turn it on. If the error message remains please contact ThermoWorks for further assistance.

Hi Lo Hi or Lo is displayed when the temperature being measured exceeds the measurement range of the thermometer.

BATTERIES

The thermometer incorporates visual low battery indication as follows:



'Battery OK'



'Battery Low': battery needs to be replaced



'Battery Exhausted': measurements may be affected

BATTERY REPLACEMENT

Power off the unit before replacing the battery.

1. Remove the screw holding the belt clip in place.

3

2. Open the battery compartment by turning the cover slightly clockwise until it stops.
3. Remove the old battery by pushing it away from the plastic clips and toward the metal hook and lifting it up.
4. Insert a new battery under the metal hook and press down until you hear a click sound.
5. Close the battery cover and replace the belt clip.

A malfunction may occur if the power is on when the battery is replaced. If a malfunction occurs, restart the device. Keep the battery away from children.

LANYARD

The lanyard can be attached to the top of the belt clip by looping the string around the point where the clip attaches to the thermometer. Feed the looped string under the belt clip and feed the remaining material through the loop in the string for a secure hold.

STORAGE & CLEANING

The thermometer should be stored at 40 to 149 °F / 4 to 65 °C.

The sensor lens is the most delicate part of the thermometer and should be kept clean at all times. Care should be taken when cleaning the lens to use only a soft cloth or cotton swab with water or medical alcohol, and allow the lens to fully dry before using the thermometer. Do not submerge any part of the thermometer.

EMC/RFI

Readings may be affected if the unit is operated within radio frequency electromagnetic field strength of approximately 3 volts per meter, but the performance of the instrument will not be permanently affected.

4

