

NIST-TRACEABLE CALIBRATION CERTIFICATE

Model Name: DishTemp

Model Number: TX-5100

Serial Number: ***Sample***

Date Placed in Service*:

Cal. Date: ***Sample***



ThermoWorks

TEST CONDITIONS

| | |
|-------------|--------|
| Temperature | 84.0°F |
| Humidity | 62% |

STANDARDS USED

| | |
|------------------|-------------------|
| Model Number | FLUKE 1521 |
| Serial Number | A7A686 |
| Manufacturer | FLUKE CORPORATION |
| Calibration Date | AUG-27-2017 |
| Calibration Due | AUG-27-2018 |

CALIBRATION DATA

| Target Temperature | Thermometer Reading | Deviation | Specification |
|--------------------|---------------------|-----------|---------------|
| 160°F | 160.3°F | 0.3°F | ±0.9°F |

* Precision temperature devices do not experience degradation or loss of accuracy between the time of calibration and the time placed into service. Consequently, the annual calibration interval begins on the date the instrument is placed into service (to be filled in by the user) rather than the calibration test date shown on the certificate.