



## CERTIFICATE OF CALIBRATION

Date of Issue: \*\*\*Sample\*\*\*  
Certificate Issued By: \*\*\*Sample\*\*\*

Instrument Code: THS-179-607  
Instrument Description: Thermopen Blue  
Instrument Serial Number: \*\*\*Sample\*\*\*

Ambient Temperature: 72°F ± 4°F, Ambient Humidity: <80 %rh

UKAS Certified Reference Instruments: Prema 6001 / Hart 1502 + SPRT

Results:	Test Temperature	Instrument Reading
	32.0°F	32.0°F
	212.0°F	212.0°F

Uncertainty of Calibration ± 0.7°F ±1 digit  
Instrument only results are conversion of electrical simulation calibration

This is to certify that the above Instrument has been calibrated against laboratory standards which are traceable via International Agreement, to all major National Standards, including the NPL and NIST.

Authorised by:   
Calibration Manager

**www.thermoworks.com**  
741 E. Utah Valley Drive, American Fork, Utah 84003