



ThermoWorks

CERTIFICATE OF CALIBRATION

Calibrated by ETI Ltd for:
ThermoWorks, Inc.
1762 W. 20 S., Unit 100
Lindon, UT 84042

Date of Issue: Wednesday, 02 October

Certificate Issued by: Rachel Cooper

Instrument Description: Thermoworks FoodCheck Thermometer

Instrument Serial No: D13400147

Ambient Temperature: 70°F ± 4°F Ambient Humidity: 40% rh ±10%

Reference Instrument: UKAS Certificate No: CN182569 & N7110582T

| Results: | Test Temperature °F | Instrument Reading °F |
|----------|---------------------|-----------------------|
| | 32.0 | 32.0 |
| | 212.0 | 212.0 |

Uncertainty of Calibration ±0.7°F

This is to certify that the above Instrument has been calibrated against laboratory Standards which are traceable, via International Agreement, to all major National Standards, including NIST.

Authorised by:

John Carswell