



**ThermoWorks**

## CERTIFICATE OF CALIBRATION

Calibrated by ETI Ltd for:  
ThermoWorks, Inc.  
1762 W. 20 S., Unit 100  
Lindon, UT 84042

Date of Issue: Wednesday, 21 August  
Certificate Issued by: Rachel Cooper  
Instrument Description: Thermoworks Therma 22 Plus Thermometer  
Instrument Serial No: D13340054

Ambient Temperature: 70°F ± 4°F      Ambient Humidity: 40% rh ±10%

Reference Instrument: UKAS Certificate No: CN182569 & N7110582T

Results:	Test Temperature °F	Instrument Reading °F
	32.0	32.0
	212.0	212.0

Uncertainty of Calibration ±0.7°F

This is to certify that the above Instrument has been calibrated against laboratory Standards which are traceable, via International Agreement, to all major National Standards, including NIST.

Authorised by:

*John Carswell*