



ThermoWorks

CERTIFICATE OF CALIBRATION

Calibrated by ETI Ltd for:
ThermoWorks, Inc.
1762 W. 20 S., Unit 100
Lindon, UT 84042

Date of Issue: Monday, 17 October
Certificate Issued by: Rachel Cooper
Instrument Description: MK2 ThermaCheck Thermometer
Instrument Serial No: D11410028

Ambient Temperature: 70°F ± 4°F. Ambient Humidity: 40% rh ±10%

Reference Instrument: UKAS Certificate No: CN102783 & N1188489T

Results:	Test Temperature °F	Instrument Reading °F
	32.0	32.0
	212.0	212.0

Uncertainty of Calibration ±0.7°F ±1 digit

This is to certify that the above Instrument has been calibrated against laboratory Standards which are traceable, via International Agreement, to all major National Standards, including NIST.

Authorised by: **John Carswell**