



CERTIFICATE OF CALIBRATION

Date of Issue: 26th July 2023

Certificate Issued By: **ML**

Instrument Code: THS-279-607

Instrument Description: Thermapen® ONE Blue

Instrument Serial Number: XXXXXXXXXX

Ambient Temperature: 72°F ± 4°F, Ambient Humidity: <80 %rh

UKAS Certified Reference Instruments: Prema 6001 / Hart 1502 + SPRT

Results:	Test Temperature	Instrument Reading
	32.0°F	32.0°F
	212.0°F	212.0°F

Uncertainty of Calibration ± 0.5°F ±1 digit

This is to certify that the above Instrument has been calibrated against laboratory standards which are traceable via International Agreement, to all major National Standards, including the NPL and NIST.

Authorised by:

Greg Hills, Calibration Manager

www.thermoworks.com

741 E. Utah Valley Drive, American Fork, Utah 84003

For recalibration, please contact ThermoWorks at calibration@thermoworks.com
or 1-800-393-6434