



CERTIFICATE OF CALIBRATION

Date of Issue: 5TH FEB

Certificate Issued By: M.W.

Instrument Code: THS-293-070

Instrument Description: ThermaData Lite logger - Black

Instrument Serial Number: D190

Ambient Temperature: 72°F ± 4°F, Ambient Humidity: <80 %rh

UKAS Certified Reference Instruments: Prema 6001 / Hart 1502 + SPRT

Results:	Test Temperature	Instrument Reading
	32.0°F	32.0°F
	122.0°F	122.0°F

Uncertainty of Calibration ± 0.9°F ± 1 digit Instrument only results are conversion of electrical simulation calibration

This is to certify that the above Instrument has been calibrated against laboratory standards which are traceable via International Agreement, to all major National Standards, including the NPL and NIST.

Authorised by:

Greg Hills, Calibration Manager

www.thermoworks.com

741 E. Utah Valley Drive, American Fork, Utah 84003