

NIST-TRACEABLE CALIBRATION CERTIFICATE



ThermoWorks

Model Name: Multi-Purpose Digital Thermometer

Model Number: PharmAlarm 1 Date Placed in Service*:

Serial Number: 1515I0000001 Cal. Date*: JAN-08

TEST CONDITIONS	
Temperature	28°C
Humidity	41%

STANDARDS USED	
Model Number	FLUKE 1521
Serial Number	A7A686
Manufacturer	FLUKE CORPORATION
Calibration Date	JUL-21
Calibration Due	

CALIBRATION DATA					
Target Temperature(°C)	Thermometer Reading		Deviation		Specification
	S1	S2	S1	S2	
-20	-20.0	-19.9	0.0	0.1	±1°C
25	25.2	25.2	0.2	0.2	±1°C
50	50.2	50.2	0.2	0.2	±1°C

* Precision temperature devices do not experience degradation or loss of accuracy between the time of calibration and the time placed into service. Consequently, the annual calibration interval begins on the date the instrument is placed into service rather than the calibration test date shown on the certificate.

ThermoWorks, Inc | 741 E. Utah Valley Dr. | American Fork, UT 84003

Ph: 801-756-7705 | Fax: 801-756-8948 | www.thermoworks.com