

NIST-TRACEABLE CALIBRATION CERTIFICATE



ThermoWorks

Model Name: MAX/MIN Recording Thermometer

Model Number: PharmAlarm 2

Serial Number: 1515I0000001 Cal. Date*: JAN-08

TEST CONDITIONS	
Temperature	28°C
Humidity	41%

STANDARDS USED	
Model Number	FLUKE 1521
Serial Number	A7A686
Manufacturer	FLUKE CORPORATION
Calibration Date	JUL-21
Calibration Due	

CALIBRATION DATA					
Target Temperature(°C)	Thermometer Reading		Deviation		Specification
	S1	S2	S1	S2	
-20	-20.0	-19.9	0.0	0.1	±1°C
25	25.2	25.2	0.2	0.2	±1°C
50	50.2	50.2	0.2	0.2	±1°C

ThermoWorks, Inc | 741 E. Utah Valley Dr. | American Fork, UT 84003

Ph: 801-756-7705 | Fax: 801-756-8948 | www.thermoworks.com