



ThermoWorks

CERTIFICATE OF CALIBRATION

Calibrated by ETI Ltd for:
ThermoWorks, Inc.
1762 W. 20 S., Unit 100
Lindon, UT 84042

Date of Issue: Monday, 03 December
Certificate Issued by: Rachel Cooper
Instrument Description: Therma 20 Thermometer
Instrument Serial No: D12480777

Ambient Temperature: 70°F ± 4°F Ambient Humidity: 40% rh ±10%

Reference Instrument: UKAS Certificate No: CN153120 & N7110582T

Results:	Test Temperature °F	Instrument Reading °F
	32	32.0
	212	212.0

Uncertainty of Calibration ±0.7°F

This is to certify that the above Instrument has been calibrated against laboratory Standards which are traceable, via International Agreement, to all major National Standards, including NIST.

Authorised by:

John Carswell