



# ThermoWorks

## CERTIFICATE OF CALIBRATION

Calibrated by ETI Ltd for:  
ThermoWorks, Inc.  
1762 W. 20 S., Unit 100  
Lindon, UT 84042

Date of Issue: Thursday, 25 April  
Certificate Issued by: Rachel Cooper  
Instrument Description: ThermaCheck Plus Thermometer  
Instrument Serial No: D13170238

Ambient Temperature:  $70^{\circ}\text{F} \pm 4^{\circ}\text{F}$       Ambient Humidity:  $40\% \text{ rh} \pm 10\%$

Reference Instrument: UKAS Certificate No: CN153120 & N7110582T

Results:	Test Temperature $^{\circ}\text{F}$	Instrument Reading $^{\circ}\text{F}$
	32.0	32.0
	212.0	212.0

Uncertainty of Calibration  $\pm 0.7^{\circ}\text{F}$

This is to certify that the above Instrument has been calibrated against laboratory Standards which are traceable, via International Agreement, to all major National Standards, including NIST.

Authorised by:

*John Carswell*